



NABL

Department of Science & Technology, India

SCOPE OF ACCREDITATION

Laboratory Hemshell Services, 503, ACME Industrial Park, W.E. Highway,
Goregaon (E), Mumbai

Accreditation Standard ISO/IEC 17025:2005

Discipline Mechanical Calibration **Issue Date** 13.04.2013

Certificate Number C-0328 **Valid Until** 12.04.2015

Last Amended on - **Page** 1 of 1

Quantity Measured/ Instrument	Range / Frequency	*Calibration Measurement Capability (\pm)	Remarks
1. FORCE UTM ¹ (Tension & Compression)	20 N to 300 kN	0.5 %	Using Load Cell as per IS 1828 Part-1, 2005
2. HARDNESS ¹ HRC & 15 N	HRC & 15 N Scale	1.5 HRC & 1.5 HR 15 N	Using Standard Block as per IS 1586:2000

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

¹ Only in Permanent Laboratory.

mallika

Convenor